



SUBBAIH INSTITUTE OF DENTAL SCIENCES SHIVAMOOGA

DEPT OF MAINTENANCE

LOG BOOK REGARDING MAINTENANCE WORKS



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 3/2/2017

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No- 2, 3 - Light Not work
2. Chair No- 1 - Switch, not working
3. _____
4. _____

Date: _____

Repair Remarks:

B.S. Kund...
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka
→ Connection regarding light is repaired
→ Suction blockage is repaired.

HOD/Section Incharge
Dept. of Oral & Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH 13, Purale, Shivamogga-577,222

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 19-4-2017.

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Leakage in dental chairs in
2. our clinic
3. Foot control is not working
4. _____

Date: 19-4-2017

Repair Remarks:

B.S. Kund...
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka
- Tubulings repaired
- Foot control spring replaced

HOD/Section Incharge
[Signature]

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 16/5/17

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chairs NO-6, 8, 9, - up & down not working
2. Chair NO-3 - handpiece connection " "
3. _____
4. _____

Date: _____

Repair Remarks:

→ Electric issue
movement of chair is repaired

Repaired by:

→ Hand piece bearings are replaced

B.S. [Signature]
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga, Karnataka

[Signature]
HOD/Section Incharge
Head of the Department
Dept. of Oral & Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 5/6/2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Gas Connection for Centric Machine in Roadway
2. _____
3. _____
4. _____

Date: 5-6-2017

B.S. [Signature]
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga, Karnataka

[Signature]
HOD/Section Incharge
Head of the Department
Dept. of Prosthodontics

Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222.

Repair Remarks:

→ Leakage check in gas line done & repaired

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 23-6-2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Cons & Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Auto clave not working
2. _____
3. _____
4. _____

Date: 23-6-2017

B.S. Subbaiah
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

[Signature]
HOD/Section Incharge

Repair Remarks:

*Autoclave faulty
coil replaced*

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 28/6/2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. HOD - room - Tubelight not working
2. _____
3. _____
4. _____

Date: 28/6/2017

For Maintenance Department Use

Repair Remarks:

Tube light was
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

Repaired by:

[Signature]
Head of the Department
Dept. of Periodontic
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577
Maintenance Incharge



**SUBBIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 22-8-2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Orthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair, 7, 8, Light is
2. not working
3. _____
4. _____

Date: 22-8-2017

[Signature]
HOD/Section Incharge

Head of the Department
Dept. of Orthodontics & Dentofacial
Orthognathics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

For Maintenance Department Use

*Bulb is replaced
with Philips, 7, 8*

Repaired by:

[Signature]
Maintenance Incharge



**SUBBIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 5-9-2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Course Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Airrotor is not working
2. X-Ray machine is not working
3. _____
4. _____

Date: 5-9-2017

B.S. Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

[Signature]
HOD/Section Incharge

Repair Remarks:

For Maintenance Department Use

*- Airrotor
- X-Ray machine screw
replaced*

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 10/9/2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgs

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No-10, 12, - Light not working
2. Chair No-2 & 3, - handpiece connector " "
3. _____
4. _____

Date: _____

B.S. Sankar
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

[Signature]
HOD/Section Incharge
Dept. of Oral
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

- Bulbs replaced
- Hand piece oiling done

Repaired by: _____

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 30-9-2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Pedodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Autoclave is not working
2. _____
3. _____
4. _____

Date: 30-9-2017

For Maintenance Department Use

B.S. Sankar
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

Repair Remarks:

- Dry coil is replaced

Repaired by: _____

[Signature]
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 21/10/2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of PROSTHODONTICS

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Light is not working in Chair No. 14.
2. Acrylizer is not working in PEP lab
3. _____
4. _____

Date: 21-10-2017

B.S. Subbaiah
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka
For Maintenance Department Use

Y. G. Reddy
HOD/Section Incharge
Dept. of Prosthetics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks: → Switching
→ coil in acrylizer repaired.

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 25-11-2017

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Course Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Light curing unit not working
2. Dental chair light is not working (in clinic)
3. _____
4. _____

Date: 25-11-2017

B.S. Subbaiah
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka
For Maintenance Department Use

[Signature]
HOD/Section Incharge
Dept. of Endodontics

Repair Remarks: - Light curing unit bulb replaced
- Dental chair bulb replaced

Repaired by:

[Signature]
Maintenance Incharge

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main, Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 07
24/04/2017

To:
SUBBAIAH INSTITUTE OF DENTAL SCIENCES
Purle
Shimoga-5772222

Line Item	Description	Quantatity	Rate	Amount
1	Filter	1	800	800.00
2	Service charges			350.00
				0.00
				0.00
				0.00
				0.00
				0.00
Subtotal				1150.00
Discount				0.00
Total				1150.00



Our Services

- Service one
- Service two

- Service one
- Service two

Purle

B.S. Subbaiah
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 3-2-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Periodontology

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No: 22 - scaler not working
2. _____
3. _____
4. _____

Date: 3-2-2018

S.G. Manjith
HOD/Section Incharge
Head of the Department
Dept. of Periodontic
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

For Maintenance Department Use

Repair Remarks: power supply cord was replaced

Repaired by:

B.S. Kund
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 20-2-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Course Endo.

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Electric pulp tester not working
2. phantom heads in pre-clinical lab not working
3. _____
4. _____

Date: 20-2-2018

B.S. Kund
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

[Signature]
HOD/Section Incharge

Repair Remarks:

- Leaking supply replacer
- phantom head service done

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 3/3/2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chairs No-1, & 3 - Not working bak
2. Chairs No- 9 & 11 - Light Not Working
3. _____
4. _____

Date: _____

Repair Remarks:

Repaired by:

Services of Shivamogga done
switches replaced

B.S. Paul
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka

D. Suresh
HOD/Section Incharge
Department of Oral and Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH.13, Purale, Shivamogga-577221

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 3/4/2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Model trimmer is not working in UG lab
2. Micro motor is not working in PEP lab.
3. _____
4. _____

Date: 3-4-2018

Repair Remarks:

Repaired by:

- model trimmer motor service done
- micro motor cord replaced

B.S. Paul
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka

[Signature]
HOD/Section Incharge
Department of Prosthodontics
Subbaiah Institute of Dental Sciences
NH.13, Purale, Shivamogga-577222.

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 9-4-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Conservator

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Airrotar's not working
2. phantom head Light is not working
3. in pre-clinical lab
4. _____

Date: 9-4-2018

B.S. Bhandari
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

[Signature]
HOD/Section Incharge
Dept. of the Department
Endodontics &
Dentistry &
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222

Repair Remarks:

→ Cartridge replaced in airrotar
→ service of phantom head done

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 10/5/2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No-2 & 3, Motorized Switches Not Working
2. Chair No-9 & 10, Light Not working
3. _____
4. _____

Date:

B.S. Bhandari
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Dr. Gantur
HOD/Section Incharge
Dept. of Oral & Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222

Repair Remarks:

→ service of motor of section
→ Bulbs replaced in chair no 9 & 10

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 23/7/2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Periodontology

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Aerotar is not working
2. _____
3. _____
4. _____

Date: 23/7/2018

S.G. Nayath
HOD/Section-Incharge

Repair Remarks: For Maintenance Department Use

Aerotar cartridge is replaced

Repaired by:

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.M. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222
[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 23-7-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Conserv Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Dental chair tubulings are
2. Leaking.
3. _____
4. _____

Date: 23-7-2023

Repair Remarks:

Tubulings replaced

Repaired by:

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.M. Road, Purle
SHIVAMOGGA-577 222, Karnataka
For Maintenance Department Use

Head of the Department
Dept. of Conservative Endodontics & Prosthetics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222
[Signature]
Maintenance Incharge

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 26-9-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Orthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair no, 9, 10, Suction
2. is not working.
3. _____
4. _____

Date: 26-9-2018

[Signature]
HOD/Section Incharge

Head of the Department
Dept. of Orthodontics & Dentofacial
Orthopedics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

For Maintenance Department Use

Repair Remarks:

*- Suction blockage
B.K. Prasad
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka*

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 6-10-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Cons. Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Air Leakage in dental
2. chair
3. suction is not working in
4. dental chair

Date: 6-10-2018

[Signature]
HOD/Section Incharge

Repair Remarks:

*→ Tubalini's
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka*
*→ Blockage in Suction
removed & repaired*

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 31-10-2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Pododontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Micromotor is not working
2. _____
3. _____
4. _____

B.S. Sankar
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Date: 31-10-2018

HOD/Section Incharge
Dept. of Pododontics

For Maintenance Department Use

Repair Remarks:

Cartridge is replaced.

Repaired by:

B.S. Sankar
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 10/11/2018

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chairs No-2 & 3, Swiches Not Working
2. Chair No-3 & 5, Had lights not "
3. _____
4. _____

Date:

B.S. Sankar
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka

HOD/Section Incharge
Dept. of Oral & Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

- Blockade of Swiches of repaired
→ Swiches of lights replaced.

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

date 27/11/2018

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To, The Maintenance Incharge
Nature/Particulars of Complaint:

1. Sand Blaster is not working
2. Model trimmer disc broken
- 3.
- 4.

Date: 27/11/2018
B.S. Kumar Principal
76 Maintenance Incharge

Repair Remarks: - Sand blaster repaired
- Model trimmer disc replaced.

Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222
Karnataka

Repaired by: Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

date 12/12/2018

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Oral Surgery

To, The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No - 465, handpiece connection Network
2. Chair No - 9616, hand light Network
- 3.
- 4.

Date: 12/12/2018
B.S. Kumar Principal
Dr. Suresh HOD Section Incharge

Repair Remarks: - Hand piece connection network repaired
- 2 bulbs are replaced.

Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222
Karnataka

Repaired by: Maintenance Incharge

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main, Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 11
23/04/2018

To:
SUBBAIAH INSTITUTE OF DENTAL SCIENCES
Purle
Shimoga-5772222

Line Item	Description	Quantity	Rate	Amount
1	Adoptor	1	400	400.00
2	Filter	1	800	800.00
3	Service Charges			350.00
				0.00
				0.00
				0.00
				0.00
Subtotal				1550.00
Discount				50.00
Total				1500.00



Our Services

- Service one
- Service two

- Service one
- Service two

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Dept. of Maintenance
COMPLAINT NOTE

Date: 17-12-2018

Name of the Department: Cong. E. H. O.

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Auto. clare is not working
2. _____
3. _____
4. _____

Date: 17-12-2018

B.S. Principal
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222
For Maintenance Department Use

[Signature]
HOD/Section Incharge

Repair Remarks:

→ Connection issue to C.O.D is repaired.

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Dept. of Maintenance
COMPLAINT NOTE

Date: 10/2/2019

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No-3 & 4, handpieces connects Not work
2. Chair No-6 & 7, head light Not Working
3. Chair No-9 & 10, Not moving up & down
4. _____

Date:

B.S. Principal
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222
For Maintenance Department Use

[Signature]
HOD/Section Incharge
Dept. of Oral & Maxillofacial Surgery
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

- cardridge
switches are charged
relaxed
Hydraulic oiling done
for movements

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 12/2/2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Micro Motor hand piece is not working.
2. _____
3. _____
4. _____

Date: 12-2-2019

B.S. [Signature]
Principal

[Signature]
HOD/Section Incharge.

Repair Remarks:

Micro Motor Cord Replaced

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 15-3-2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Conservative Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Bleaching unit not working.
2. _____
3. _____
4. _____

Date: 15-3-2019

B.S. [Signature]
Principal

[Signature]
HOD/Section Incharge

Repair Remarks:

power supply fault is rectified.

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 21/A/2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No-4 & 5, Hand light Not working
2. Chair No-9 & 10, Not moving up & down
3. Chair No-2, Suction Not working
4. _____

Date: _____
B.S. Kundu
Principal
 For Maintenance Department Use
 Subbaiah Institute of Dental Sciences
 NH-13, H.H. Road, Purale, Shivamogga-577222, Karnataka

[Signature]
 HOD/Section Incharge
 Dept. of Oral & Maxillofacial Surgery
 Subbaiah Institute of Dental Sciences
 NH-13, Purale, Shivamogga-577222

Repair Remarks: Bulbs are replaced
→ Electric issues repaired
→ Service of suction motor done

Repaired by: _____

[Signature]
 Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 6-5-2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Pedodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Airrotor is not working
2. _____
3. _____
4. _____

Date: 6-5-2019

Repair Remarks: _____

Repaired by: _____

Lubrication done
[Signature]
Principal
 Subbaiah Institute of Dental Sciences
 NH-13, H.H. Road, Purale, Shivamogga-577222, Karnataka

[Signature]
 Head of the Department
 Dept. of Pedodontics
 HOD/Section Incharge

[Signature]
 Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 30-5-2019

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Orthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. chair no 4, 5, ^{no} UP & down
2. movements is
3. _____
4. _____

Date: 30-5-2019

[Signature]
HOD/Section Incharge
Head of the Department

Repair Remarks: For Maintenance Department Use

Dept. of Orthodontics & Dentofacial
Orthopaedics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

*Ecc, PCB, replaced with
chain.*
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale

[Signature]
Maintenance Incharge

Repaired by:



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 26-7-2019

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Convey Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Water leakage from dental
2. chairs.
3. _____
4. _____

Date: 26-7-2019

B.S. [Signature]
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577222

[Signature]
HOD/Section Incharge
Head of the Department

Repair Remarks:

- connectors are replaced.

[Signature]
Maintenance Incharge

Repaired by:



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 9-8-2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Department of Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair no: 16 - Injection is not working
2. _____
3. _____
4. _____

Date: 9-8-2019

For Maintenance Department Use

HOD/Section Incharge
Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

suction motor was serviced

Repaired by:

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale,
SHIVAMOGGA-577222

Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 3/9/2019

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No 4 is not functioning properly
2. Cyrax is not working
3. Auto clean water leakage is there
4. _____

Date: 3-9-2019

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale,
SHIVAMOGGA-577222

HOD/Section Incharge

Repair Remarks:

→ movement is not working
→ Autoclave heating coil is replaced
→ Ceyser coil is replaced.

Repaired by:

Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 10/9/2019

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Oral Surgery

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair No-3 & 4, Pt. trolley Not moving
2. Chair No-6 & 7, Head light Not working
3. Chair No-2, Suction Not Working
4. _____

Date: _____

B.S. Kundu
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

D. Brantner
HOD/Section Incharge
Department
Oral Surgery
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222

Repair Remarks:

- 7 trolley wheels are replaced
- Electric issues related to bulbs repaired
- Suction Line blockage cleared.

Repaired by: _____

Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 19-11-2019

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Cons & Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Airrotor is not working
2. _____
3. _____
4. _____

Date: 19-11-2019

B.S. Kundu
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

[Signature]
HOD/Section Incharge
Department
Cons & Endo
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222

Repair Remarks:

Cartridge of airrotor is replaced.

Repaired by: _____

Maintenance Incharge

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main, Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 10
26/04/2019

To:
SUBBAIAH INSTITUTE OF DENTAL SCIENCES
Purle
Shimoga-5772222

Line Item	Description	Quantity	Rate	Amount
1	Filter	1	800	800.00
2	Service charges			350.00
				0.00
				0.00
				0.00
				0.00
				0.00
				0.00
			Subtotal	1150.00
			Discount	0.00
			Total	1150.00

Our Services

- Service one
- Service two

- Service one
- Service two



Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 26-2-2020

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Pedodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chaire no- 3, 4 is not working
2. i.e no movement
3. _____
4. _____

Date: 26-2-2020

HOD/Section Incharge
B.S. Subrament
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Repair Remarks:

Repaired by:

For Maintenance Department Use
→ PCB chair no 3 & 4 replaced
→ ECC chair replaced.
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 3-3-2020

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: COR & Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Leakage in Pressure
2. lines in clinic
3. _____
4. _____

Date: 3-3-2020

Repair Remarks:

Repaired by:

For Maintenance Department Use
Leakage in Pressure
line Identified &
repaired.
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

HOD/Section Incharge
[Signature]
Principal
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222, Karnataka



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 16/3/2020

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. vibrator is not working in ceramic lab
2. vacuum mixer is not working in ceramic lab
3. _____
4. _____

Date: 16-3-2020

B.S. [Signature]
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale,
Shivamogga-577222, Karnataka

[Signature]
HOD/Section Incharge
Head of the Department
Dept. of Prosthodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks: power supply issue with vibrator is repaired
switch is vacuum mixer is repaired

[Signature]
Maintenance Incharge

Repaired by:



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 19-3-2020

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Orthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Auto claver is not working
2. _____
3. _____
4. _____

Date: 19-3-2020

[Signature]
Principal
For Maintenance Department Use
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale,
Shivamogga-577222, Karnataka

[Signature]
Head of the Department
Dept. of Orthodontics & Dental
Orthopaedics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 25-4-2020

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. HOD ROOM - Refrigeration Leakage
2. _____
3. _____
4. _____

Date: 25-4-2020

[Signature]
HOD/Section Incharge

For Maintenance Department Use

Repair Remarks:

Refrigerator was sent for service. Properly given another refrigerator.

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 17/8/2020

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Model trimmer water is not coming
2. airrotar is not working.
3. Chair No 6 airrotar connector is not working
4. _____

Date: 17-8-2020

B.S. [Signature]
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

[Signature]
HOD/Section Incharge

Repair Remarks:

- 1 → Coil in model trimmer repaired
- 2 → Cartridge in airrotar replaced
- 3 → Airrotar connectors are replaced.

Repaired by: *[Signature]* Maintenance Incharge

Head of the Department
Dept. of Prosthodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main, Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 08
10/04/2020

To:

SUBBAIAH INSTITUTE OF DENTAL SCIENCES

Purle

Shimoga-5772222

Line Item	Description	Quantity	Rate	Amount
1	Adoptor	1	400	400.00
2	Filter	2	800	1600.00
3	Service Charges			350.00
				0.00
				0.00
				0.00
				0.00
			Subtotal	2350.00
			Discount	50.00
			Total	2300.00



Our Services

- Service one
- Service two

- Service one
- Service two

[Handwritten signature]

Principal

Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 11-11-2020

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair no-14 - light is not working
2. _____
3. _____
4. _____

Date: 11-11-2021

For Maintenance Department Use

Repair Remarks:

switch is replaced

Principal

Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Repaired by:

[Signature]
HOD/Section Incharge
Head of the Department
Dept. of Periodontic
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222
[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 19/02/2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Micro motor in puller not working
2. lathe machine stone broken
3. _____
4. _____

Date: 19/02/2021

For Maintenance Department Use

Repair Remarks:

Micro-motor and replaced

Wire in Lathe machine is replaced

Repaired by:

[Signature]
HOD/Section Incharge
Head of the Department
Dept. of Prosthodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222
[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 11-11-2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Congu Endo

To, The Maintenance Incharge
Nature/Particulars of Complaint:

1. Suction not working
2. in dental chair
3. _____
4. _____

Date: 11-11-2021

B.S. Reddy
Principal

[Signature]
Incharge

Repair Remarks:

- Suction
blockage removed
repaired.

Repaired by:

Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 28-11-2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Periodontics

To, The Maintenance Incharge
Nature/Particulars of Complaint:

1. Implant section - Autoclave - not working
2. _____
3. _____
4. _____

Date: 28/11/2021

For Maintenance Department Use

HOD Section Incharge

Repair Remarks:

implant section of autoclave changed

[Signature]
Principal

Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repaired by:

Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 21-7-2021

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: orthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Electro Solder is not working
2. _____
3. _____
4. _____

Date: 21-7-2021

B.S. Prasad
HOD/Section Incharge
Head of the Department

Repair Remarks:

For Maintenance Department Use

Dept. of Orthodontics & Dentofacial
Orthopaedics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Transformer is replaced
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 30-7-2021

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: pedodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. chair no 8 is Leaking
2. _____
3. _____
4. _____

Date: 30-7-2021

B.S. Prasad
Head of the Department
Dept. of Pedodontics & Preventive
HOD/Section Incharge

Repair Remarks:

For Maintenance Department Use

Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Chair no 8 Light is replaced
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale
SHIVAMOGGA-577 222, Karnataka

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 30/9/2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Acralizer is not working in UG clinic
2. Hydraulic press is not working " "
3. chair no 10 spittoon water is not coming
4. Refrinator is not working

Date: 30/9/2021

B.S. Subudhi
Principal

Y. G. Reddy
HOD/Section Incharge

Repair Remarks: → Acrylizer repaired
→ Changing of Hydraulic Press is done
→ Pipeline of spittoon is changed (no 10)
→ Refrinator is repaired.

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 18-10-2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Conve Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Lights are not working
2. in dental chairs.
3. _____
4. _____

Date: 18-10-2021

B.S. Subudhi
Principal

[Signature]
HOD/Section Incharge

Repair Remarks:

Power light source repaired.

Repaired by:

[Signature]
Maintenance Incharge

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main. Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 06
20/04/2021

To:
SUBBAIAH INSTITUTE OF DENTAL SCIENCES
Purle
Shimoga-5772222

Line Item	Description	Quantity	Rate	Amount
1	Adaptor	1	400	400.00
2	Filter	1	800	800.00
3	Service Charges			350.00
				0.00
				0.00
				0.00
				0.00
Subtotal				1550.00
Discount				50.00
Total				1500.00

 Our Services

- Service one
- Service two
- Service one
- Service two




Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 6-12-2021

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Chair no - 2 - section is not working.
2. _____
3. _____
4. _____

Date: 6-12-2021

[Signature]
HOD/Section Incharge

For Maintenance Department Use

Repair Remarks:

Suction connector
changed

Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental
NH-13, Purale, Shivamogga-577 222

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 16-2-2022

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Cons Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Light cure unit is not
2. working.
3. _____
4. _____

Date: 16-2-2022

For Maintenance Department Use

Repair Remarks:

Power supply
replaced.

Repaired by:

[Signature]
Maintenance Incharge

Head of the Department
Dept. of Endodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577 222



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 19/4/2022

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Dept of Prosthodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Micromotor is not working
2. Heating unit is not working
3. Chain no 12, 16, 20 light is not working
4. Plug in geyser is not working

Date: 19-4-2022

For Maintenance Department Use
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purale, Shivamogga-577222, Karnataka

Head of the Department
Section Incharge
Dept. of Prosthodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks: micromotor coil
- Heating coil in heating unit is
- Bulbs in chain no 12, 16, 20, changed

Repaired by: plug in geyser is replaced

Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 5-9-22

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Reader room - Fan is not working
2. _____
3. _____
4. _____

Date: 5-9-22

For Maintenance Department Use

HOD/Section Incharge
Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222

Repair Remarks:

capacitor of fan changed

Repaired by:

Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 15-10-2022

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Dental Model Trimmer is
2. not working
3. _____
4. _____

Date: 15-10-2022

B.S. Subudhi
Head of the Department
Dept. of Periodontics & Perio-implant
HOD/Section Incharge

For Maintenance Department Use

Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222.

Repair Remarks:

Capacitor is replaced.
Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Repaired by:

[Signature]
Maintenance Incharge



SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 11-11-2022

Dept. of Maintenance
COMPLAINT NOTE

Name of the Department: Core & Endo

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. changing of tubulings of
2. dental chair
3. _____
4. _____

Date: 11-11-2022

For Maintenance Department Use

[Signature]
Head of the Department
HOD/Section Incharge

Repair Remarks:

old & Leaking tubulings
replaced.

Repaired by:

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 16-11-22

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Periodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. Seminar room - LED Display is not working.
2. _____
3. _____
4. _____

Date: 16-11-22

S. G. Nayth
HOD/Section Incharge

For Maintenance Department Use

Head of the Department
Dept. of Periodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222.

Repair Remarks: power socket in wire changed.

Repaired by: B. A. Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

[Signature]
Maintenance Incharge



**SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222**

Date: 29/12/2022

**Dept. of Maintenance
COMPLAINT NOTE**

Name of the Department: Dept of Prostodontics

To,
The Maintenance Incharge
Nature/Particulars of Complaint:

1. High Speed lathe is not working in Ceramic lab.
2. Burn out furnace is not working in Ceramic lab.
3. Model frame disc is broken.
4. _____

Date: 29/12/2022

B. A. Principal
Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577222, Karnataka

[Signature]
HOD/Section Incharge
Dept. of Prostodontics
Subbaiah Institute of Dental Sciences
NH-13, Purale, Shivamogga-577222.

Repair Remarks: → winding of Lathe motor done.
→ Heating coil of furnace is changed.
→ Disc of Model Trimmer is changed.

Repaired by:

[Signature]
Maintenance Incharge

Aqua pearl

American technology water purifiers

Karibasappa Building, Opp Pruthvi Mansion 100 Feet Road, Main, Road, Vinoba Nagara,
Shivamogga, Karnataka 577201

Invoice

Invoice No: 22
08/04/2022

To:
SUBBAIAH INSTITUTE OF DENTAL SCIENCES
Purle
Shimoga-5772222

Line Item	Description	Quantity	Rate	Amount
1	Filter	1	800	800.00
2	Service charges			350.00
				0.00
				0.00
				0.00
				0.00
				0.00
			Subtotal	1150.00
			Discount	0.00
			Total	1150.00



Our Services

- Service one
- Service two

- Service one
- Service two

Principal

Subbaiah Institute of Dental Sciences
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

SUBBAIAH INSTITUTE OF DENTAL SCIENCES,
NH-13, PURALE, SHIVAMOGGA-577222

Date: 23/3/23

COMPLAINT NOTE

From the Department: conservative dentistry (Room no-4)

To, The Maintenance Incharge (IT Incharge)
Nature/Particulars of Complaint: Mr. Ganesh

- ① - To, Install Team viewer, Any desk & Google chrome software to RVC system
- ② - To give net connection to RVC computer

4. _____
Head of the Department

Dept. of Conservative Dentistry (9739442258)

Subbaiah Institute of Dental Science

NH-13, Purale, Shivamogga-577222

For Maintenance Department Use only

Principal Signature
Subbaiah Institute of Dental Science
NH-13, H.H. Road, Purle
SHIVAMOGGA-577 222, Karnataka

Date: 23/3/23

Repair Remarks:

Team viewer,
Any desk, google chrome
software installed
Internet connection provided

Repaired by:

Received signature

Maintenance Incharge Signature

MAINTENACE WORK OF PHATOM HEAD LAB



MAINTENACE WORK OF DENTAL CHAIRS IN CONSERVATIVE DENTISTRY DEPARTMENT



MAINTENACE WORK OF AUTOCLAVE

MAINTENACE WORK OF MAIN AIR COMPRESSOR OF THE COLLEG



MAINTENACE WORK OF RO DRINKING WATER UNIT OF THE COLLEG

ROUTINE CLEANING OF DEPARTMENTS

